

BLANK PAGE



Indian Standard

ENGINEERING METROLOGY — CYLINDRICAL SQUARES — SPECIFICATION

(First Revision)

भारतीय मानक

इंजीनियरी माप विज्ञान — बेलनाकार गुनिया — विशिष्टि (पहला पुनरीक्षण)

UDC 531.747.1

© BIS 1990

BUREAU OF INDIAN STANDARDS MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG NEW DELHI 110002

FOREWORD

This Indian Standard (First Revision) was adopted by the Bureau of Indian Standards on 13 March 1990, after the draft finalized by the Engineering Metrology Sectional Committee had been approved by the Light Mechanical Engineering Division Council.

This Indian Standard was originally published in 1973. This first revision has been brought out in the light of prevailing technical practices followed in India.

In the preparation of this standard, considerable assistance has been derived from BS 939: 1977 'Specification for engineers squares (including cylindrical and block squares)' issued by the British Standards Institution (BSI).

Specific requirements for handle have been incorporated. The requirements relating to marking of angular positions 0°, 90°, 180° and 270° adjacent to each end face has been deleted, as it does not affect the performance of cylindrical square.

Indian Standard

ENGINEERING METROLOGY—CYLINDRICAL SQUARES—SPECIFICATION

(First Revision)

1 SCOPE

1.1 This standard covers the requirements of cylindrical squares of sizes up to 750 mm length.

2 REFERENCES

2.1 The following Indian Standards are necessary adjuncts to this standard:

IS No.

Title

2473:1975

Dimensions for centre holes (first revision)

7018 (Part 1): 1983

Technical supply conditions for gauges: Part 1 General

(first revision)

3 TERMINOLOGY

3.1 For the purpose of this standard, working surfaces mean both end faces and cylindrical surface.

4 TYPES

4.1 The cylindrical squares shall be of Type A and Type B (see Table 1).

5 DIMENSIONS AND TOLERANCES

5.1 Dimensions and tolerances for cylindrical squares shall be as given in Table 1.

6 MATERIAL

6.1 The cylindrical squares shall be made of suitable quality steel such as T215Cr12 and T105CrlMn60 as specified in IS 7018 (Part 1): 1983.

7 HARDNESS

7.1 The hardness of the working surfaces shall conform to IS 7018 (Part 1): 1983.

8 GENERAL REQUIREMENTS

- **8.1** In order to reduce weight, it is recommended that cylindrical squares of 300 mm length and larger should be of hollow section.
- **8.2** All cylindrical squares shall have their outer edges rounded off to radius of 1.5 mm or chamfered to 1 mm \times 45°.

- **8.3** The end faces of cylindrical squares shall be relieved to a depth of 2 mm.
- 8.4 Cylindrical squares shall be provided with permanent conical centres conforming to Type B Centre (with relief) of IS 2473: 1975 for purposes of manufacture and reconditioning.
- 8.5 It is recommended that cylindrical squares particularly of 300 mm length and above shall be provided with suitable lifting handles. In such case, provision for fixing arrangement of handle shall be provided on both sides in the relieved surfaces to facilitate working on both end faces of cylindrical squares.

9 DESIGNATION

9.1 The cylindrical square shall be designated by type, length and IS No., for example the cylindrical square of Type B and length 300 mm shall be designated as:

Cylindrical Square Type B 300 IS 6952.


10 MARKING

- 10.1 Each cylindrical square shall be marked legibly and permanently and marking shall not protrude above the working surfaces of the cylindrical square. Each cylindrical square shall be marked with following particulars:
 - a) Length,
 - b) Type,
 - c) Indication of the source of manufacture, and
 - d) Number of this standard.
- 10.2 The cylindrical squares may also be marked with the Standard Mark, details of which may be obtained from Bureau of Indian Standards.

11 PRESERVATION AND PACKING

- 11.1 The working surfaces of cylindrical squares shall be protected against climatic conditions by suitable non-corrosive treatment.
- 11.2 Each cylindrical square shall be packed in a suitable wooden box provided for the purpose.

Table 1 Dimensions and Tolerances for Cylindrical Square (Clauses 4 and 5)



Туре	Nominal Size	D	d	x	y	z
	mm	mm	mm	μm	μm	μm
,	.75	50	30	1.2	1	1
A	150	70	50	2	1.5	1.2
	220	90	70	3	2	2
	300	95	70	5	2.2	3
В	450	120	85	7	4	4.2
	600	140	100	9	5	6
	750	160	115	11	6,	7:5

Standard Mark

The use of the Standard Mark is governed by the provisions of the Bureau of Indian Standards Act, 1986 and the Rules and Regulations made thereunder. The Standard Mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well defined system of inspection, testing and quality control which is devised and supervised by BIS and operated by the producer. Standard marked products are also continuously checked by BIS for conformity to that standard as a further safeguard. Details of conditions under which a licence for the use of the Standard Mark may be granted to manufacturers or producers may be obtained from the Bureau of Indian Standards.

Bureau of Indian Standards

BIS is a statutory institution established under the Bureau of Indian Standards Act, 1986 to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Director (Publications), BIS.

Revision of Indian Standards

Indian Standards are reviewed periodically and revised, when necessary and amendments, if any, are issued from time to time. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition. Comments on this Indian Standard may be sent to BIS giving the following reference:

Doc: No. LMD 05 (4924)

Amendments Issued Since Publication

Date of Issue	Text Affected	
,		
	Date of issue	

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002 Telephones: 331 01 31, 331 13 75

Telegrams: Manaksanstha (Common to all Offices)

Regional Offices:

Central	ntral: Manak Bhavan, 9 Bahadur Shah Zafar Marg NEW DELHI 110002				
_				25 22 22	

Eastern: 1/14 C.I.T. Scheme VII M, V.I.P. Road, Maniktola (CALCUTTA 700054) 37 86 62

Northern: SCO 445-446, Sector 35-C, CHANDIGARH 160036 2 18 43

Southern: C.I.T. Campus, IV Cross Road, MADRAS 600113 41 29 16

Western: Manakalaya, E9 MIDC, Marol, Andheri (East) 6 32 92 95 BOMBAY 400093

Branches: AHMADABAD. BANGALORE. BHOPAL. BHUBANESHWAR. COIMBATORE. FARIDABAD. GHAZIABAD. GUWAHATI.

HYDERABAD. JAIPUR. KANPUR. PATNA. THIRUVANANTHAPURAM.